



"Your Agent to Product Certifications and Licenses Throughout the World"

RTCA/DO-160F

TEST CERTIFICATION

Manufacturer: CareFusion

Models: LTV 800, LTV 900, LTV 950, LTV 1000, LTV1100, LTV 1150, LTV1200, LTMII, SprintPack

Test Date: February 17, 2010

Issue Date: February 15, 2012

Test Report #170210

RTCA/DO-160F Sections 20, 21 Certification

The LTV Ventilator System has passed full compliance with the requirements of RTCA/DO-160F Section 20, Category S, and Section 21, Category M. Testing was performed using a Battery Power Source for the Ventilator and LTMII. Such certification is required for use aboard all commercial airlines landing and/or taking off within the United States. This certification complies with the Air Carriers Act and the Department of Transportation's Nondiscrimination on the Basis of Disability in Air Travel policy.

Approved by:

Duane R. Bagdons
President
International Certification Services Inc.
1100 Falcon Ave.
Glencoe, Minnesota 55336
(320) 864-4444

Customer Representative:

Gregory Smith
CareFusion
Infusion Respiratory Services
17400 Medina Road Suite 100
Minneapolis, Minnesota 55447-1341