



VELA™ ventilator

**NOTE:** The VELA $^{\bowtie}$  ventilator quick tips card is not intended as a replacement for the operator's manual.

You must become completely familiar with the operator's manual before using the VELA ventilator.

# VELA noninvasive positive-pressure ventilation (NPPV) modes

- NPPV A/C NPPV with assist control ventilation
- NPPV/SIMV NPPV with synchronized intermittent mandatory ventilation
- NPPV CPAP PSV NPPV and continuous positive airway pressure (demand breath) with pressure support ventilation

## **Primary breath controls**

Displayed control	Description	Range	
Rate	Breaths per minute	2 to 80 BPM	
Vt	Tidal volume	50 to 2,000 mL	
Insp Pres	Inspiratory pressure	1 to 100 cmH <sub>2</sub> O	
Peak Flow	Peak inspiratory flow	10 to 140 L/min	
Insp Time	Inspiratory time	0.3 to 10 sec	
Insp Pause	Inspiratory pause	OFF, 0.1 to 2 sec	
PSV	Pressure support	OFF, 1 to 60 cmH <sub>2</sub> O	
PEEP	Positive end-expiratory pressure	0 to 35 cmH <sub>2</sub> O	
Flow Trig	Inspiratory flow trigger	1 to 20 L/min	
*O <sub>2</sub>	Percentage of oxygen*	21 to 100%	
NPPV Pinsp	NPPV inspiratory pressure	0 to 40 cmH <sub>2</sub> O	
NPPV PSV	NPPV pressure support ventilation	OFF, 1 to 40 cmH <sub>2</sub> O	

<sup>\*</sup>Used only when connected to a high pressure oxygen source

### Advanced breath controls in NPPV modes

- PC Flow Cycle sets the percentage of peak inspiratory flow that terminates the inspiratory phase of a PC breath.
- PSV Cycle sets the percentage of peak inspiratory flow that terminates the inspiratory phase of a PSV breath.
- Bias Flow sets the continuous flow through the patient breathing circuit.
- PSV Tmax controls the maximum inspiratory time of a pressure-supported breath.

Advanced control	Range	Primary control	NPPV A/C	NPPV SIMV	NPPV CPAP PSV
PC Flow Cycle	5 to 70%	Insp Time	•	•	
PSV Cycle	5 to 70%	NPPV PSV		•	•
Bias Flow	10 to 20 LPM	Flow Trig	•	•	•
PSV Tmax	0.3 to 3 sec	NPPV PSV		•	•

# **Leak compensation**

In NPPV modes, the default setting is on. Leak compensation is up to 40 LPM, in addition to the bias flow (10 to 20 LPM, default 10 LPM).

**VELA NPPV CPAP PSV set-up** – for traditional noninvasive bilevel ventilation

To set up **VELA NPPV CPAP PSV**:

- 1. Press the MODE bar to access the MODE SELECT window.
- 2. Press NPPV CPAP PSV.



#### **VELA NPPV CPAP PSV set-up** (continued)

- 3. Set primary breath controls parameters:
  - Set the PEEP to the desired pressure value. This setting is equivalent to CPAP/EPAP pressure.



b. Set the NPPV PSV to the desired pressure value. This setting is the pressure value plus the set PEEP value that equals the ordered peak inspiratory pressure (Ppeak) (e.g., if PEEP = 5 cmH<sub>2</sub>O and NPPV PSV = 5 cmH<sub>2</sub>O, then Ppeak = 10 cmH<sub>2</sub>O).



- **c.** Set the  $O_2$  and flow trigger levels.
- 4. Set the advanced settings:
  - **a.** Press **ADV SETTING** to access the **ADVANCED SETTINGS** window.
  - **b.** Press **NPPV PSV** to access the PSV (flow) Cycle and PSV Tmax controls.
  - c. Set the PSV Cycle. Generally, set it at 45 to 50% to facilitate the transition to exhalation.
  - d. Set the PSV Tmax. A setting between 1 and 1.5 seconds provides backup termination if large leaks occur during the PSV Cycle.

#### **VELA NPPV CPAP PSV set-up** (continued)



- With NPPV CPAP PSV, the APNEA MODE backup settings only appear in the MODE SELECT window. Set the APNEA MODE backup:
  - a. Press Apnea Settings.
  - **b.** In **APNEA MODE**, press Volume or Pressure.
  - c. Set the primary breath controls for the APNEA MODE backup.
- 6. Set the alarms:
  - a. Press LIMITS to access the **ALARM LIMITS** window.
  - **b.** Press the desired alarm, and using the data dial, change the setting.
  - c. To accept the new alarm setting, press ACCEPT. To restore the previous setting, press CANCEL.
  - d. Press LIMITS to exit the ALARM LIMITS window.



7. Press Mode Accept to begin ventilation with the new settings.

## **VELA** monitoring options

#### The three screens that enable monitoring include:

- Main displays 5 user-definable parameters plus 3 waveforms.
- Monitors displays 15 user-definable parameters.
- Trends displays up to 24 hours of monitored data, 4 histograms and 11 monitored values—all user defined.

To select the desired screen, press the **NPPV A/C** screen to access the **SCREEN SELECT** window.



## Patient interface/mask

#### Specific to NPPV and a dual limb circuit:

- · Has no exhalation ports—Is not a mask designed for CPAP or OSA
- Features labeling for noninvasive ventilation use.
- Eliminates ported masks or "whisper valves," as leakage is not necessary for proper operation.

#### Mask fit

#### To ensure proper fit:

- Use the mask sizing template, which usually comes with the mask, to ensure the correct size.
- Do not overtighten the mask or headgear, as that can lead to pressure areas.
- Ensure that the facial contours are optimized to prevent leaks (e.g., dentures are in place, and facial hair is minimal).



## GLOBAL HEADQUARTERS

Vyaire Medical, Inc. 26125 North Riverwoods Blvd Mettawa, IL 60045 USA

# vyaire.com

© 2020 Vyaire. Vyaire, the Vyaire logo and all other trademarks or registered trademarks are property of Vyaire Medical, Inc., or one of its affiliates. Please read the complete Instructions For Use that come with the devices or follow the instructions on the product labelling. VYR-GBL-2000164